



The meteorological measuring set (MMS) AN/TMQ-41, NSN 6660-01-386-3906, uses two types of light fixtures. One fixture takes a halogen bulb and the other an incandescent bulb.

The MMS AN/TMQ-41A uses one type of light fixture—an incandescent bulb.

Some MMS AN/TMQ-41 units contain a light fixture that requires a dual contact, 28V, halogen bulb. This bulb needs to be local purchased from the manufacturer, Ushio America Inc., using CAGE 0H3N2 and a part number of 1000917. Their phone number is 800-838-7446.

Other MMS AN/TMQ-41s and all AN/TMQ-41As contain a light fixture that requires a single contact, 28V, incandescent bulb, NSN 6240-00-155-7965.

REMEMBER, THE INCANDESCENT BULB IS **NOT INTERCHANGEABLE** WITH THE HALOGEN BULB!

"PUT THE INCANDESCENT WHERE THE HALOGEN BELONGS AND YOU'LL **RUIN THE LIGHT.**"



When replacing the light fixture in any MMS, use the one that uses the incandescent bulb. The one that uses the halogen is being phased out. The incandescent bulb fixture needs to be ordered by part number S2312-201 and CAGE 78366 on a DD Form 1348-6.