

Local TMDE Sends PATS for Calibration

LOOKS LIKE
YOU DON'T
SEND ME FOR
CALIBRATION,
BUT YOUR
LOCAL TMDE
DOES.

GOOD!
THAT'LL
BE EASIER
FOR ME.



Dear Editor,

On Pages 41-42 of PS 716 (July 12), you told CBRN specialists to send their M41 PATS (protection assessment test system) to the TMDE activities at either Pirmasens Army Depot or Redstone Arsenal for calibration.

Actually, the local TMDE support centers should be doing that, not the individual CBRN specialists. The CBRN specialists should turn in their PATS for calibration every 12 or 18 months to the local TMDE facility, which will then ship PATS to the Pirmasens or Redstone addresses you listed. Determine your calibration interval by checking TMDE's TB 43-180, *Calibration and Repair Requirements for the Maintenance of Army Materiel*. Direct questions to the local TMDE labs.

Bob Herndon
US Army TMDE Activity
Regions West

Editor's note: Got it! Thanks
for straightening us out.

M26 Decon...

HOW DO YOU GET HOSE FILTERS?



Dear Half-Mast,

TM 3-4230-238-10 says we can clean the filter for the M26 decon's metering valve suction hose. But eventually you need to replace the filter and I can't find any mention in the TM of a filter NSN. Does it have one?

W.O.

Dear Mr. W.O.,

We're in harmony on this one.
Order the filter with NSN 4730-
12-324-0555.



↑
Filter

Half-Mast